



Certificate of Calibration.

Serial Number of Gauge	66888/1
Size and Type	100mm Direct Mounting Glycerine Filled
Dial Marking	0 – 300lbs PSI / 20 Bar
Tested Against	Deadweight Tester 15878/380H (P786)
Temperature of Test	20°C
Accuracy of Determination	BS: EN – 837 – 1 Class 1

The Gauge was tapped before taking a reading. Positive corrections are to be added to and negative corrections subtracted from the Gauge reading.

(The figures given under “Pressure Falling” were taken during reduction of pressure from the maximum scale marking).

True Pressure (lbs)	Gauge Readings (lbs)			
	Pressure Rising	Correction	Pressure Falling	Correction
0	0	NIL	0	NIL
50	49	+1	49	+1
100	100	NIL	100	NIL
150	150	NIL	150	NIL
200	200	NIL	200	NIL
250	250	NIL	250	NIL
300	300	NIL	300	NIL

Date: 31st July 2009

Signed: 

Inspector

**All CALIBRATION EQUIPMENT
USED IS TRACEABLE TO
NATIONAL STANDARDS**

Unit 6, Abbey Park Industrial Estate, Abbey Road, Barking
Essex IG11 7BT